



## **Manufacture's FAA acceptance criteria Declaration of Conformity**

Precision Medical Inc. has determined the following POC devices conform to all applicable FAA acceptance criteria for POC carriage and use on board aircraft.

EasyPulse POC, PM4150 Series

EasyPulse POC, PM4130 Series

All the above devices have tested and passed the requirements of;  
RTCA/DO-160F, Environmental Conditions and Test Procedures for Airborne Equipment.  
Section 21.5, Radiated RF Emissions, Category M

Test reports are available upon request to;

**Precision Medical Inc.**

300 Held Drive

Northampton Pa 18067

[www.precisionmedical.com](http://www.precisionmedical.com)

(610) 262-6090